Installation Instructions
In-Line Vacuum Exhaust Muffler
Part #SVA95785

Description:
In the event that a vacuum system must be installed in an area where space is limited, (such as under a sink) it may be difficult to mount a separator or recycler properly. The lack of a separator or recycler may result in an increase in discharge noise. To remedy this situation, Apollo Dental Products offers a compact exhaust muffler which may be installed directly in the discharge line and will significantly reduce the amount of discharge noise.

Installation:
1. Be sure that the vacuum pump is off.
2. Locate the muffler as close to the sewer inlet as possible and place directly in-line with vacuum discharge.

   NOTE: Be sure to note the direction of flow through the muffler.



Cabinet Installation Diagram

Warranty Information: 90 Days
All ADP accessories are thoroughly inspected and tested in accordance with rigid specifications and standards. Our accessories are guaranteed against any defective material and workmanship from the date of shipment; provided that the installation, operation, and maintenance is done in accordance with ADP procedures as outlined in our Installation and Maintenance Guides. Warranty cards must be returned to ADP within ten days of installation to effect warranty. No other warranties or guarantees, expressed or implied are made.

ADP's obligation under the warranty is to provide parts for the repair or, at its option, to provide the replacement product (excluding labor). All special, incidental and/or consequential damages are excluded. We will not issue credit for accessories without first attempting to correct the problem in the field. Written notice of breach of warranty must be given to ADP within the warranty period. The warranty does not cove damage resulting form improper installation or maintenance, accident or misuse. The warranty does not cover damage resulting from the use of cleaning, disinfecting or sterilizing chemicals and processes. The warranty does not cover vacuum failures due to hard water deposits. Failure to follow instructions provided in ADP's Installation and Maintenance Guides may void the warranty.

Apollo Dental Products, Inc. • 245 W. Dakota Ave. • Clovis, CA 93612 • Technical Support
(559) 292-1444 • (800) 233-4151 • Fax (559) 292-1555 • www.apollodental.com

AMI60251 Rev. 1/01